LIFEPAK CR PLUS® DEFIBRILLATOR
USER'S CHECKLIST

Unit Serial Number _____

Department/Location _____

Instruction	Recommended Corrective Action	Date / Initials							
1. Check readiness for:									
OK indicator	None.								
CHARGE-PAK indicator	Replace CHARGE-PAK™ and QUIK-PAK™ electrode packet.								
ATTENTION indicator	Refer to operating instructions.								
WRENCH indicator	Contact authorized service personnel.								
2. Check Use By date on QUIK-PAK™ electrode packet.	Replace electrode packet and CHARGE-PAK™ if date passed.								
3. Check additional supplies.	Replenish as needed.								
4. Check defibrillator for:									
Damage or cracks	Contact authorized AED coordinator								
Foreign substances	Clean the device.								
5. Other:									